

購 買 (英文) 規 格 書

COMMODITY DESCRIPTION

규격서 B

關稅分類番號 H. S. NO.	品目番號 ITEM NO.	品 目 및 規 格 DESCRIPTION	單 位 Unit	數 量 Quantity
9018.50-1010		Retina Angiograph (Model : Spectralis HRA)	System	1

A. Feature

1. Confocal and digital laser scanning system for indocyanine green angiography
2. Simultaneous acquisition of FAG and ICG
3. Outstanding quality and contrast of early and early and late phase images
4. Acquisition of single images or temporal sequences or focal series
5. Digital image processing and archiving

B. Specifications

1. Scan method : Confocal Laser Scanning
2. Source of capture images : 795nm infrared laser
3. Field of view : 15×15, 20×20, 30×30 degrees
wide angle with ultra wide field lens. 102 degrees
4. Digitalized image size : 768×768 at 15°×15°
1024×1024 at 20°×20°
1536×1536 at 30°×30°
5. Scan time per image : 0.032sec.
6. Image repetition frequency : 20Hz
7. Digital resolution : 5 / pixel (for all field)
8. Multiple image display : 70 images simultaneously
9. 3D information : separate imaging of choroidal and
retinal vascular system

C. Consisting of

Ophthalmic Equipment

Model:Spectralis HRA

1 Sys

Consisting of :

- Laser Scanning Camera
- Operation Panel with Foot-switch
- Adjustable Camera Mount Table
- Parallelogram Camera Mount
- Camera Mount Base Plate
- Chin Rest
- Power Supply
- Cable for Operation Panel to Power Supply
- Cable for Digital I/O Board to Power Supply
- Operation Software on CD
- Software Protection Connector

D. Remark

Warranty is 3years.

안 과 장 김 상 덕 (인)

